



DAILY BED HUDDLE

June 3, 2013

Privileged and Confidential: For purposes of self-critical analysis

Date	Bed Flow Issues	Patient Issues	Safety Issues	Resolution	Responsible Party	Incident Report
6/3/2013	Census: 247 ED holding: 1 tele observation. SDS holding: 0 PACU holding: 0 Critical Care: CT ICU-11 available beds ICU-0 available beds CCU-2 available bed 6 East-1 available telemetry bed Med surg: 7 West-10 available beds 7 East-3 available beds 6 West-0 available beds CCL: 4 cases (1 OPD)					
6/3/2013			A problem in the OR Room 7 was referred to facilities and biomed over the weekend when an alarm continued to go off.			

			The source of the problem was a power cord to the OR table. The cord was replaced.			
--	--	--	--	--	--	--